
Switch mode wall plug in lithium ion chargers are designed to effectively charge lithium ion batteries while protecting the batteries from over charging. All charger ratings are based on a nominal input of 90-264VAC/47-63Hz. The output current ranges from 500mA to 1.3A.



## Features:

- Medical approved (EN 60601, versions UL 60601)
- Wall plug in form with exchangeable AC plugs
- 3 step charge control
- Switch mode charger with universal input voltage
- Charges 1 to 7 cell packs
- Protected against reverse polarity and short circuit
- Current detection as charge termination
- Custom specifications on request



Category	Specification			
Model number	452241L Series			
Input rating	Nominal 90-264VAC/47-63Hz			
Cell count	1-7 cells			
Switch frequency	≈ 40kHz			
Leakage current from battery	≈ 0			
Operating temperature	-25°C - +40°C			
Ripple	< 100 mV p-p			
Insulation class	Class II			
Insulation voltage	4000VAC (primary) / 5640VDC (secondary)			
Electrical safety approvals	UL 60601-1(1,2,3 cell versions certified), EN 60950, EN 60601-1, EN 60335-2-29			
EMC standards	EN60601-1-2 (Medical), EN 61000-6-3 (Emission), EN 61000-6-1 (Immunity)			
Reliability (MIL-GDBK-217F)	MTBF > 250.000 hours at 30°C and full load			
Input connection	2 pin IEC 320-C7 (input cordset not included)			
Output connection	2.5mm x 5.5mm x 9mm barrel plug, center positive*			
Dimensions/weight	4.06" x 1.81" x 1.54" (103 x 46 x 39mm) / .291 lbs. (130g)			



## **Functionality:**

Once the charger is attached to a lithium ion battery and then plugged in, the charging process will begin. The charger will subject the lithium ion battery to three steps of charging.

During the first step, the charger enters into 'fast charge' mode. During this step, the charger is in constant current mode with the current rate remaining at its maximum rating.

In step two, the battery will be nominally 80-95% charged, and the charger switches into constant voltage mode (current is no longer being provided at its maximum rate). The charger will continue to provide a constant voltage until the charge current decreases to the charge termination level.

In the third step, the charge process is complete. Charging has ceased as no current flows to the battery.


#### **Versions:**

Version			Charge control				
Model number	Cell count	Max output power (W)	Efficiency 100% at 230V	Charge current	Charge voltage	Charge termination I<	Charge resumes with Vbat< or mains turned on
452241-LA	1	5.5	58%	1.3A +5/-7%	4.2V +.05V	100mA +/- 25%	4.05 +.05/15V
452241-LB	2	10.9	72%	1.3A +5/-7%	8.4V +/1V	100mA +/- 25%	8.1V +.1/15V
452241-LC	3	15.1	77%	1.2A +5/-7%	12.6V +/1V	100mA +/- 25%	12.3V + .1/15V
452241-LD	4	15.1	80%	0.9A +5/-7%	16.8V +/1V	100mA +/- 25%	16.4V +.1/15V
452241-LE	5	14.7	82%	0.7A +5/-7%	21.0V +/1V	100mA +/- 25%	20.5V +.1/15V
452241-LF	6	15.1	82%	0.6A +5/-7%	25.2V +/1V	100mA +/- 25%	24.6V +.1/15V
452241-LG	7	16.4	82%	0.56A +5/-7%	29.4V +/1V	100mA +/- 25%	28.7V +.1/15V


#### LED status indicator:

The lithium ion charger has an LED status indicator to inform the user of its status. During step 1 and step 2, the LED will appear red in color. During step 3, the LED will appear green in color.

# Charge curve:



<sup>\*</sup>When specifying product, please consult with Cell-Con to verify that the specifications identified on this data sheet are current.



Removeable AC Plugs

